



City of Milwaukee-Public Health Laboratories
841 North Broadway, Room 205 Milwaukee, WI 53202-3653
Phone Number: (414)286-3526 Fax Number: (414)286-5098
Autoreporting Lab

Submitter copy to:

HOME ENVIRONMENTAL HEALTH LEAD PROGRAM
841 N Broadway Ave
1st Floor
Milwaukee, WI 53202

Order ID : Z4270006
LRN : 0000324981
Auxiliary ID : 20830
Date Collected: 06/26/25
Date Received : 06/27/25

Requested by: [REDACTED]

FINAL

Patient Name: CP/N/MARYLANDAV, 2418

DOB:

Age:

Sex:

Ord. Comm: 20830, HEH - 4 Soil samples

- 1) Play area, Side D by exit 5
- 2) Play area, Side A by exit 2
- 3) Raised Garden, Side B by dripline
- 4) Garden soil (ground level) Side B

CHEMISTRY

TEST-NAME	RESULT	AB	NRML-RANGE	UNITS	DATE-TIME
COLLECTED 06/26/25 14:40 RECEIVED 06/27/25 15:39					

Lead in Soil:

Sample 1	*22			mg Pb/kg	07/01/25 15:30
wt = 0.50738 g					
Sample 2	*45			mg Pb/kg	07/01/25 15:30
wt = 0.50193 g					
Sample 3	*94			mg Pb/kg	07/01/25 15:30
wt = 0.50274 g					
Sample 4	*23			mg Pb/kg	07/01/25 15:30
wt = 0.50299 g					
Test Method	*see below				06/27/25 16:04
Sample Preparation: Modified ASTM E1726 per PbSOP					
Analytical Method: Modified EPA Method 7000B per PbSOP					
Reporting Limit: 8.5 mg Pb/kg					

Sample results have not been corrected for field blank or analytical blank. Results related only to those samples tested. All sample information is provided to the lab by the client unless otherwise stated.

QC results associated with these samples were acceptable
continued on next page

Legends: L-Low, H-High, AB-Abnormal, P-Panic, C-Critical, X-Extreme



City of Milwaukee-Public Health Laboratories
 841 North Broadway, Room 205 Milwaukee, WI 53202-3653
 Phone Number: (414)286-3526 Fax Number: (414)286-5098
 Autoreporting Lab

Submitter copy to:

HOME ENVIRONMENTAL HEALTH LEAD PROGRAM
 841 N Broadway Ave
 1st Floor
 Milwaukee, WI 53202

Order ID : Z4270006
 LRN : 0000324981
 Auxiliary ID : 20830
 Date Collected: 06/26/25
 Date Received : 06/27/25

Requested by: [REDACTED]

FINAL

Patient Name: CP/N/MARYLANDAV, 2418

DOB:

Age:

Sex:

Ord. Comm: 20830, HEH - 4 Soil samples

- 1) Play area, Side D by exit 5
- 2) Play area, Side A by exit 2
- 3) Raised Garden, Side B by dripline
- 4) Garden soil (ground level) Side B

continued

CHEMISTRY

TEST-NAME	RESULT	AB	NRML-RANGE	UNITS	DATE-TIME
unless otherwise noted.					
Data reviewed and approved by the QA Coordinator/Technical Manager.					
Accrediting body: AIHA-LAP, LLC; Lab ID #102186.					

Legends: L-Low, H-High, AB-Abnormal, P-Panic, C-Critical, X-Extreme